

REPORT DOCUMENTATION PAGEForm Approved
OMB No. 0704-0188

Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing this collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden to Department of Defense, Washington Headquarters Services, Directorate for Information Operations and Reports (0704-0188), 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302. Respondents should be aware that notwithstanding any other provision of law, no person shall be subject to any penalty for failing to comply with a collection of information if it does not display a currently valid OMB control number. PLEASE DO NOT RETURN YOUR FORM TO THE ABOVE ADDRESS.

1. REPORT DATE (DD-MM-YYYY)		2. REPORT TYPE Technical Papers		3. DATES COVERED (From - To)	
4. TITLE AND SUBTITLE				5a. CONTRACT NUMBER	
				5b. GRANT NUMBER	
				5c. PROGRAM ELEMENT NUMBER	
6. AUTHOR(S)				5d. PROJECT NUMBER 3DS8	
				5e. TASK NUMBER RF9A	
				5f. WORK UNIT NUMBER	
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) Air Force Research Laboratory (AFMC) AFRL/PRS 5 Pollux Drive Edwards AFB CA 93524-7048				8. PERFORMING ORGANIZATION REPORT	
9. SPONSORING / MONITORING AGENCY NAME(S) AND ADDRESS(ES) Air Force Research Laboratory (AFMC) AFRL/PRS 5 Pollux Drive Edwards AFB CA 93524-7048				10. SPONSOR/MONITOR'S ACRONYM(S)	
				11. SPONSOR/MONITOR'S NUMBER(S)	
12. DISTRIBUTION / AVAILABILITY STATEMENT Approved for public release; distribution unlimited.					
13. SUPPLEMENTARY NOTES					
14. ABSTRACT <div style="text-align: center; font-size: 2em;">20030103 132</div>					
15. SUBJECT TERMS					
16. SECURITY CLASSIFICATION OF:			17. LIMITATION OF ABSTRACT A	18. NUMBER OF PAGES	19a. NAME OF RESPONSIBLE PERSON Leilani Richardson
a. REPORT Unclassified	b. ABSTRACT Unclassified	c. THIS PAGE Unclassified			19b. TELEPHONE NUMBER (include area code) (661) 275-5015

Standard Form 298 (Rev. 8-98)
Prescribed by ANSI Std. Z39.18

1 item enclosed

⑤ Paper Rec'd After 30-day Deadline = 23 days until Deadline } No mch
issued

MEMORANDUM FOR PRS (In-House Publication)

FROM: PROI (STINFO)

06 Nov 2002

SUBJECT: Authorization for Release of Technical Information, Control Number: **AFRL-PR-ED-AB-2002-269**
P.A. Strakey (PRSA), "Assessment of Multiple Scattering Errors of Laser Diffraction Instruments"
(abstract only)

Int'l Conf. on Liquid Atomization & Spray Systems
(Sorrento, Italy, 13-18 July 2003) (Deadline: 29 Nov 02)

(Statement A)

Peter
55673

Doug (LPM)
56174

Assessment of Multiple Scattering Errors of Laser Diffraction Instruments

P.A. Strakey

Air Force Research Laboratory | AFRL/PRSA
10 E. Saturn Blvd., Edwards AFB, CA 93524

Characterization of high-pressure rocket injector sprays often involves using commercial laser-based diagnostic instruments to measure droplet size in an environment well beyond the intended range of the instruments. Even in cold-flow measurements, this environment often involves very large number densities of droplets ($N > 10^4 \text{ cc}^{-1}$) as well as refractive index gradients in the surrounding gas. As a result, optical attenuation and multiple scattering of the probe beam(s) and scattering signal is often the limiting factor in the ability to extract useable data.

Several experimental studies on multiple scattering errors have been reported previously using optically thick suspensions of particles. Most of these studies were limited to relatively small particles ($< 150 \mu\text{m}$) and either monodispersed or narrow particle size distributions. Hamidi and Swithenbank investigated bi-modal size distributions of glass beads with a maximum size of $105 \mu\text{m}$.¹ They found significant errors in measured mean size and proposed an analytical correction factor that could be applied to two-parameter size distributions. Gulder carried out a similar study reporting multiple scattering effects on Sauter mean diameter.² He proposed a correction scheme using a model-independent analysis that was validated for D_{32} up to $100 \mu\text{m}$. Dodge simulated a dense spray environment using a linear array of nominally identical atomizers with a maximum D_{32} of about $20 \mu\text{m}$.³ His findings showed that the correction scheme developed by Hamidi and Swithenbank greatly reduced the error in D_{32} at transmissions as low as 5%, but did not fully compensate for the effects of multiple scattering. More recently, Harvill *et al.* reported results for suspensions of aluminum oxide particles with a volume median diameter of $50 \mu\text{m}$.⁴ They found that without a correction, the error in measured volume median diameter was as much as 40% at a transmission of 5%. They also demonstrated the validity of a statistically based correction scheme that did an excellent job of correcting for multiple scattering effects.

Liquid bi-propellant rocket injectors often generate large, non-spherical droplets as a result of the relatively low injection velocities and high chamber back-pressures which create high deformation stresses on the droplet. The goal of this study is to assess the capability and limitations of the laser diffraction technique in dense sprays typical of rocket operating conditions. This involves a relatively broad range of sizes from tens of microns to nearly a millimeter in diameter.

Experimental Setup

In this study, two commercial laser-diffraction instruments were tested. The first was a Malvern 2600c Master Sizer system and the second was a Malvern SprayTech instrument. In order to assess the accuracy and limitations of the two instruments, it was necessary to simulate the dense spray environment with a two-phase medium of known particle number density and distribution. This was accomplished using a dispersion of solid, spherical polystyrene microspheres and distilled water in a magnetically stirred glass test cell.

Separate experiments were conducted with each instrument using both monodispersed microspheres at concentrations ranging from 90% to 1% transmission and sizes ranging from $30 \mu\text{m}$ to $650 \mu\text{m}$. Experiments were also conducted with polydispersed mixtures of beads over the same range of concentrations and sizes. The polydispersed mixtures consisted of six different bead sizes in relative concentrations that approximated lognormal size distributions typical of large-scale rocket injectors.

Results and Discussion

Results from the polydispersed experiments are presented as a percent error in the volume-weighted volume mean diameter, D_{43} from the actual size. D_{43} was chosen as a representative indicator of accuracy

because both instruments use a process of inverting the light scattering data to obtain a particle volume distribution and are thus geared toward providing maximum accuracy in a volume mean diameter. Figure 1 contains plots of measurement error, expressed as a percentage of D_{43} as a function of transmission for each instrument for the polydispersed bead mixtures. In both cases, the multiple scattering correction option has been used. The Spraytech instrument (Fig. 1(a)) showed good accuracy over the range of transmissions studied here. The instrument was accurate to within $\pm 10\%$ in the transmission range of 90% to 2%, which is the stated lower range of the instrument. The instrument produced reasonably good results even at a transmission of 1%. The minimum error occurred at about 60% transmission for all the distributions measured.

The results from the 2600c instrument (Fig. 1(b)) were not as encouraging. Even with the multiple scattering correction, the measured D_{43} began to drop significantly below a transmission of about 10%. At a transmission of 2% the measured D_{43} was found to be less than the actual D_{43} by as much as 45%. This is consistent with the fundamental problem of multiple scattering in which the overall angle of light scattering increases with each scattering event. Since scattering angle is inversely proportional to particle size, the increase in overall scattering angle results in a smaller measured size. The multiple scattering correction algorithm used by this instrument did not fully compensate for this effect, however, it did greatly improve the accuracy over using no correction at all. This result is consistent with the findings of Dodge who found that even with the correction, the instrument under-estimated mean size by as much as 20% at a transmission of about 5%.

The disparity in Fig. 1 can be explained by the differences in the multiple scattering correction schemes employed by the two instruments. The 2600c uses a relatively simple analytical correction model while the algorithm used in the Spraytech instrument is based on a statistical approach. The statistical approach requires fewer assumptions at the expense of increased computational complexity.

In terms of dense spray measurements, the limiting factor for the 2600c instrument appears to be the multiple scattering correction which is reliable only to about 10% transmission for the conditions studied here. For the SprayTech instrument, the accuracy is limited by signal-to-noise errors that become significant below about 2% transmission. The instrument is still useable however down to a transmission of about 1%.

References

1. Hamidi, A. A. and Swithenbank, J., "Treatment of Multiple Scattering of Light in Laser Diffraction Measurement Techniques in Dense Sprays and Particle Fields", *Journal of the Institute of Energy*, No. 101, June, 1986.
2. Gulder, Omer, "Multiple Scattering Effects in Dense Spray Sizing by Laser Diffraction", *Aerosol Science and Technology*, Vol. 12, 1990, pp. 570-577.
3. Dodge, L. G., "Change of Calibration of Diffraction-based Particle Sizers in Dense Sprays", *Optical Engineering*, Vol. 23, No. 5, 1984, pp. 626-630.
4. Harvill, T. L., Hoog, J. H. and Holve, D. J., "In-Process Particle Size Distribution Measurements and Control", *Particle and Particle Characterization*, Vol. 12, 1995, pp. 309-313.




Figure 1: Percent error in measured D_{43} verses transmission for (a) SprayTech and (b) 2600c instruments.